




Patient: Jane Doe
DOB: 01/01/1902
Gender: Female



Provider: Albert C Domm
Practice: Lifetime Sciences Test Account
4037 Rural Plains Circle,
Suite 150
Franklin, TN, 37064



Specimen: Swab - Nasopharyngeal
Identifier: SP000-063-923
Collected: 06/02/2022 02:18pm
Received: 06/02/2022 02:28pm
Reported: 06/03/2022 10:42am



Targets Not Detected			
Target	Type	Result	
Influenza A	Viral	Not Detected	
Influenza B	Viral	Not Detected	
SARS-CoV-19	Viral	Not Detected	

Methodology

Total nucleic acid extraction is performed using DNA MagMax MVP11 Kits, which are validated by their manufacturer to yield isolated nucleic acids of sufficient quantity and quality from the relevant sample types. Presence of pathogenic nucleic acids was determined by conducting RT-PCR reactions with primers designed to target species-specific (or marker-specific) genomic regions. All sample runs contain a minimum of one negative extraction control (NEC), one negative template control (NTC), and one positive control of known pathogenic composition (PTC). All RT-PCR reactions are conducted using TaqMan chemistry from Thermo Fisher. Nucleic acid-based pathogen detection is performed on the QuantStudio 12K platform. This platform was designed by Applied Biosystems, Inc. and utilizes quantitative real-time PCR in conjunction with fluorescently-labelled nucleic acid probes. A DETECTED result signifies that amplification of genus, species, or marker-specific (dependent on the analyte) genetic markers was observed, based on validated detection metrics.

Disclaimer and Limitations

Nucleic acid-based pathogen detection (PD) is a laboratory-developed test (LDT) and as such, is not cleared by the FDA. Nucleic acid-based pathogen detection analysis is intended to aid physicians in identifying underlying pathogens within a patient sample to help advise on possible treatment avenues. It should not be used in the contexts of diagnosis or supplant physician recommendations. These results should be interpreted along with clinical presentations and/or other laboratory results. These results and their clinical interpretations may not be accurate if the patient information/sample supplied is incomplete or inaccurate. If these results do not match clinical presentations, additional testing is recommended. The LDTs reported here were developed and had its performance metrics determined by Lifetime Sciences. Lifetime Sciences is certified under the Clinical Laboratory Improvements Act (CLIA) to perform high-complexity clinical laboratory testing. For other questions/concerns, please contact Lifetime Sciences.